




FORMAT FOR STORM MAGNETOGRAMS  
(SAMPLE ONLY)



SEE CALIBRATION DATA FOR SCALE VALUES AND BASELINE VALUES

NORMAL MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF
D	1. Aug.	31. Aug.	1.45' /mm	1.0' E	
H	1. Aug.	31. Aug.	5.28 nT/mm	11201 nT	+8.5 nT/C <sup>o</sup>
Z	1. Aug.	31. Aug.	6.50 nT/mm		-2.8 nT/C <sup>o</sup>
	1. Aug.	23. Aug.		51510 nT	
	24. Aug.	31. Aug.		51500 nT	

STORM MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF
D	1. Aug.	31. Aug.	3.0' /mm	52.0' W	
H	1. Aug.	31. Aug.	22.2 nT/mm	11136 nT	-0.8 nT/C <sup>o</sup>
Z	1. Aug.	31. Aug.	24.0 nT/mm	51206 nT	-3.7 nT/C <sup>o</sup>

PRELIMINARY QUIET MEAN ABSOLUTE VALUES		
D	H	Z
39.0' E	11355 nT	51590 nT
Computed from following days: 8. 9. 11. 16.		