




FORMAT FOR STORM MAGNETOGRAMS
(SAMPLE ONLY)



SEE CALIBRATION DATA FOR SCALE VALUES AND BASELINE VALUES

NORMAL MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Jan.	31. Jan	1.45' /mm	1.0' E	
H	1. Jan.	31. Jan.	5.28 nT/mm	11202 nT	+8.5 nT/C ^o
Z	1. Jan.	13. Jan.	6.50 nT/mm	51523 nT	-2.8 nT/C ^o
	14. Jan.	31. Jan.	"	51515 nT	"

STORM MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Jan.	31. Jan.	3.0' /mm	52.0' W	
H	1. Jan.	31. Jan.	22.2 nT/mm	11140 nT	-0.8 nT/C ^o
Z	1. Jan.	31. Jan.	24.0 nT/mm	51188 nT	-3.7 nT/C ^o

PRELIMINARY QUIET MEAN ABSOLUTE VALUES		
D	H	Z
34.5' E	11375 nT	51575 nT
Computed from following days: <u>12. 18. 19. 20. 21.</u>		