




FORMAT FOR STORM MAGNETOGRAMS
(SAMPLE ONLY)



SEE CALIBRATION DATA FOR SCALE VALUES AND BASELINE VALUES

CALIBRATION DATA FOR:

JULY

1981

NORMAL MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. July	31. July	1.45' /mm	1.0' E	
H	1. July	31. July	5.28 nT/mm	11200 nT	+8.5 nT/C ^o
Z	1. July	31. July	6.50 nT/mm		-2.8 nT/C ^o
	1. July	24. July		51500 nT	
	25. July	31. July		51510 nT	

STORM MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. July	31. July	3.0' /mm	52.0' W	
H	1. July	31. July	22.2 nT/mm	11134 nT	-0.8 nT/C ^o
Z	1. July	31. July	24.0 nT/mm	51206 nT	-3.7 nT/C ^o

PRELIMINARY QUIET MEAN ABSOLUTE VALUES		
D	H	Z
39.0' E	11367 nT	51570 nT
Computed from following days: 10. 15. 16. 21. 29.		