




FORMAT FOR STORM MAGNETOGRAMS
(SAMPLE ONLY)



SEE CALIBRATION DATA FOR SCALE VALUES AND BASELINE VALUES

CALIBRATION DATA FOR: MARCH 1981

NORMAL MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Mar.	31. Mar	1.45' /mm	1.0' E	
H	1. Mar.	31. Mar.	5.28 nT/mm	11203 nT	+8.5 nT/C ^o
Z	1. Mar.	31. Mar.	6.50 nT/mm	51485 nT	-2.8 nT/C ^o

STORM MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Mar.	31. Mar.	3.0' /mm	51.7' W	
H	1. Mar.	31. Mar.	22.2 nT/mm	11138 nT	-0.8 nT/C ^o
Z	1. Mar.	31. Mar.	24.0 nT/mm	51188 nT	-3.7 nT/C ^o

PRELIMINARY QUIET MEAN ABSOLUTE VALUES		
D	H	Z
36.5' E	11365 nT	51575 nT
Computed from following days:		
<u>9. 22. 23. 24.</u>		