


FORMAT FOR STORM MAGNETOGRAMS
(SAMPLE ONLY)



CALIBRATION DATA FOR: SEPTEMBER 1981

NORMAL MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Sep.	30. Sep.	1.45' /mm	1.0' E	
H	1. Sep.	30. Sep.	5.28 nT/mm	11201 nT	+8.5 nT/C ^o
Z	1. Sep.	30. Sep.	6.50 nT/mm	51500 nT	-2.8 nT/C ^o

STORM MAGNETOGRAPH					
COMPONENT	PERIOD		CALIBRATION		
	FROM	TO	SCALE VALUE	BASELINE	TEMP. COEFF.
D	1. Sep.	30. Sep.	3.0' /mm	52.0' W	
H	1. Sep.	30. Sep.	22.2 nT/mm	11136 nT	-0.8 nT/C ^o
Z	1. Sep.	30. Sep.	24.0 nT/mm	51206 nT	-3.7 nT/C ^o

PRELIMINARY QUIET MEAN ABSOLUTE VALUES		
D	H	Z
39.4' E	11353 nT	51580 nT
Computed from following days: <u>1. 8. 17. 21. 23.</u>		